Arrangement for Visualizing Molecules

The invention relates to a method for visualizing molecules, interactions between molecules and molecular processes in a sample by using the single dye labeling method, as well as arrangements for carrying out such methods.

The object of highly sensitive detection systems is the observation on the level of individual atoms or molecules, respectively. This has first been made possible by the invention of the "Scanning Probe"-microscopy methods (EP 0 027 517-B1; Binnig et al., Phys. Rev. Lett. 56 (1986), pp. 930-933; Drake et al., Science 243 (1989), pp. 1586-1589). Yet, the detection of single molecules has also been made possible by optical methods. The effective conversion of light by fluorescent molecules also allowed for the detection of individual fluorophores in liquids by confocal fluorescence microscopy as well as for effecting a high resolution spectroscopy of single dye molecules at low temperatures.

The first real imaging of single dye molecules by optical means was achieved by near field optical scanning microscopy (Betzig et al., Science 262 (1993), 1422-1425). With this method, a spatial resolution of about 14 nm was achieved, which is far below the optical diffraction limit, yet application of this method is limited to immobile objects.

Furthermore, it has been possible to image single fluorescence-labeled myosin molecules on immobilized actin filaments by conventional microscopy and illumination times of seconds (Funatsu et al., Nature 374 (1995), pp. 555-559). This method is limited to observations in the immediate proximity of the substrate surface (distance of up to about 100 nm).

In GB 2 231 958, the characterization of the fluorescence of solid specimens by time resolved fluorescence spectroscopy is described. In doing so, not even the single molecule sensitivity is achieved so that a detection of single fluorophores is not described. In this instance, the fluorescence is fixed in the specimen and immobile. Analyzed areas in the specimen are not subjected to microscopy, but scanned by a focus in the scanning method.

In principle, the method described in US 5,528,046 is suitable for detecting single fluorophores, yet only if they have been fixed in clusters on surfaces. This measurement in the dry state (not in the aqueous phase) is, of course, not suitable for biological preparations because the functional and structural integrity of the biological preparations is destroyed by the process of drying. The apparatus constituting a prerequisite for the method described in US 5,528,046 thus is not suitable for observing single molecules in biological samples. Moreover, also a shifting of the sample which is

coupled with the detection and analysis arrangement, is not provided. Accordingly, with the methodology used there, in principle it is not possible to provide an image of biomolecules which must take place within a few milliseconds (50 milliseconds at the most), since with the device described in US 5,528,046, the illumination time is around 60 seconds.

According to US 4,793,705 it is, as such, maintained that individual particles or molecules can be identified, yet in fact this method proved to be impossible to be carried out, since individual fluorescence molecules could not be detected clearly and much less could be imaged. The ratio of signal to background of the individual observation being approximately 0.2 was extremely low so that fluctuation of the background was approximately of equal size as the signal. Also by the consecutive repetitions of the observation as well as by the parallel collection by two detectors this is not changed, either. Thus, also this method is not applicable to single molecule detection in solution or in biological systems. The method is not an imaging microscopy, but merely accumulates spatial information in sequence. Moreover, the control of a relative movement by the detection and analysis device is missing.

Single molecule detection by means of fluorescence spectroscopy in large volumes are described in DE 197 18 016 A and US 5,815,262 A, as well as sequen-

tial fluorophore detection in the confocal scanning method (WO 97/43611). Yet also with these systems, the spatial microscopy and the temporal observation of single molecule movements, particularly in biological systems (e.g. in cells) are not possible.

For allowing biological systems to be analyzed in their complete extent and for their natural function and for their physiological mode of action, visualization of individual fluorophores in complex systems and in movement as simultaneously as possible is required, i.e. real imaging microscopy (no scanning of a focus) with single molecule sensitivity, without restriction to the immediate vicinity to the sample surface or to the substrate surface. So far, the movement of single dye molecules has merely been illustrated for fluorescence-labeled lipids in an artificial lipid membrane system (Schmidt et al., PNAS 93 (1996), pp. 2926-2929). The methodology used for this has generally been termed "single dye tracing" (SDT) method, since with this it is possible to trace the path of a single fluorescence-labeled molecule and of several ones simultaneously exactly and (as a single molecule) stoichiometrically without requiring an interaction (amplification) with other components (e.g. by binding, spatial close relationship etc.) for signal emission.

Mapping of the positions and tracing of the movements of single dye labeled molecules in cellular sys-

tems which would be required for a study of molecules or interactions between molecules in live systems is, however, not possible with the methods described. On the one hand, this is due to the fact that, in contrast to flat (planar) artificial lipid membranes, live cells are three-dimensional so that molecular movements in general do not occur in an optical image plane, and, on the other hand, to the fact that cells always have a certain autofluorescence which may interfere with the fluorescence microscopy-visualizing procedure proper. Moreoever, it has been considered impossible so far to analyze a plurality of such cellular systems with a suitable detection and analyzing method so rapidly that both the resolution in the single-molecular range is maintained and also molecular movements of the mole-cules to be detected can be observed.

Primarily the pharmaceutical industry is more and more interested in methods with which a high throughput screening (HTS) of a large number of possible test molecules is possible. Particularly for HTS methods, however, the hitherto described methods for SDT are not suitable.

Thus, the object of the present invention consists in modifying the SDT method such that screening, in particular HTS, is made feasible therewith.

Moreover, an SDT method is to be provided by which molecular processes of one or several different type(s)

- 5 -

of molecules, preferably also in cellular systems, can be pursued in their real space-time dimension, wherein information on colocalization of molecules as well as on the stoichiometry of molecular associates and conformations of the molecules are also to be obtained.

Moreover, an arrangement and a method are to be provided, by means of which the imaging of fluorescence-labeled molecules in their distribution over entire biological systems, in particular cells, is made possible. Furthermore, imaging of consequences of molecular movements and processes is to be made feasible so that a three-dimensional image, with time resolution, of complex biological systems, such as cells, is made possible.

According to the invention, this object is achieved by an arrangement for visualizing molecules, their movements, and interactions between molecules, and molecular processes in a sample, in particular molecules and processes in biological cells, by using the single dye tracing (SDT) method, comprising

• at least one source of light for large-area fluorescence excitation via single or multi-photon absorption by equal or different marker molecules on molecules in the sample,

• a sample holding means for accommodating the sample,

• a highly-sensitive detection and analysis system comprising a charged coupled device (CCD) camera, the

sample or the sample holding means, respectively,
and/or the detection and analysis system being shift-
able relative to each other during the measuring pro-
cess, and

• a control unit for coordinating and synchronizing il-
  lumination times and, optionally, wave lengths, lat-
  eral or vertical movement of the sample or of the
  sample holding means, respectively, with the sample,
  as well as, optionally, the positioning and shifting
  of the images of each sample position of the pixel
  array of the CCD camera.

  Due to the large-area fluorescence excitation,

preferably 100 to 10,000 μm^2, depending on the applica-
tion, imaging of the excited molecules in a large re-
gion may be very rapid and may be read into the pixel
array of the CCD camera. In doing so, only the source
of light needs to be suitable for large-area fluores-
cence excitation. Here, a preferred source of light is
a laser. Preferably an argon laser, a dye laser and/or
a two-photon fluorescence excitation laser is used,
with acousto-optical switching between these sources of
light and for temporal sequence of the illumination.

  The CCD camera to be used according to the inven-
tion preferably comprises a frame shift mode and a con-
tinuous readout mode.

  According to the invention, preferably a CCD cam-
era is used which comprises one or several of the fol-

lowing properties: it is N_2-cooled; it has a large

pixel array, in particular a pixel array $\geq$1340 x 1300

pixels; it is capable of making a conversion from pho-

tons into electrons of 0.8 to 0.9 in the optical range;

it has a readout noise of merely a few electrons per

pixel, preferably of merely 0 to 10, in particular 3

to 7, electrons per pixel, at 1 µs/pixel readout rate;

and/or it has a lineshift rate of $>3\times10^5$/s.

    With the arrangement of the invention, a relative

movement between the sample and the detection or analy-

sis system, respectively, is necessary, which relative

movement may be continuous or step-wise. Preferably,

the lateral movement shall be possible to be continu-

ously constant, and the vertical movement shall be at-

tained by step-wise shifting of the focussing plane.

    The control unit of the arrangement according to

the invention serves to coordinate and synchronize the

illumination times and – if several wave lengths are

used – to control the wave lengths, and also to coordi-

nate the lateral or vertical relative movements between

sample and detection and analysis system. Such control

may, e.g., be effected by the CCD camera itself or by

an arrangement comprising a pulse transmitter and a

software for controlling the source(s) of light and the

(relative) movement of the sample. In this instance,

preferably, the control unit can also coordinate and

synchronize the positioning and the shifting of the im-

ages to each sample position on the pixel array of the CCD camera and control and coordinate the readout and the evaluation of the pixel array images.

The arrangement according to the invention preferably comprises an epifluorescence microscope, in particular an epifluorescence microscope with a collecting efficiency of fluorescence quantums as electrons in pixels of the CCD camera of >3%, at a 40- to 100-fold magnification.

As the sample, the arrangement according to the invention preferably comprises a molecule library prepared by combinatorial chemistry.

It is more preferred that the sample comprises a multi-well plate or a micro (nano) titer plate.

Primarily if an epifluorescence microscope having a parallel beam region is used as source of light, preferably an galvano-optic mirror is provided in the parallel beam region, with which, e.g., an even faster data storage is enabled than is provided by the readout rate or frame transfer, respectively, of the CCD camera.

In the system according to the invention, "dyed" single molecules (e.g. fluorescence-labeled biomolecules) of a sample, in particular of a biological sample which is provided on a sample holding means, can be imaged on the pixel array of the CCD camera by the highly sensitive detection and analysis system, it be-

ing possible to continuously and constantly shift the sample and/or the detection and analysis system relative to each other. For such relative movement, the frame shift of the CCD camera may be used so that the signals (e.g. the fluorescence photons) of each single molecule, after conversion into electrons ("counts") will be collected in the same pixels until the single molecule signal (number of "counts") exceeds a certain minimum signal/noise ratio (which ensures the significance of the measurement).

With the arrangement according to the invention a decisive progress has been achieved over the afore-mentioned methods for detecting single molecules in artificial lipid membranes (Schmidt et al., Laser und Optoelektronik 29(1) (1997), pp. 56-62), in that the system used there can also be operated as HTS method with the arrangement of the invention, on account of the shifting procedure, and, therebeyond, can be simply used on complete biological cells. By enlarging the highly sensitive detection and analysis system with a scanning system, suprisingly, a constant single molecule sensitivity could be maintained in a simple manner (since each CCD camera in principle has a frame shift (the shifting and readout speed from line to line of the pixel array of the camera)), with a maximized throughput rate, and fluorophores on or in complete cells could be imaged within a very short period of

time (approximately in 120 ms).

The high-resolution detection and analysis system according to the invention must be suitable for imaging the sample on the sample holding means insofar as it must have a pixel array image of the sample with a localization of individual molecules of at least 50 to 100 nm. To this end, according to the invention, a charged coupled device camera (CCD camera) is used which hitherto has already been particularly suitable in epifluorescence microscopy. With this, precisions of the localization of less than 30 nm can be attained without any problem.

When collecting the data, the lateral movement of the sample preferably should be carried out constantly and continuously, since an abrupt stopping or a high acceleration of the sample may cause the molecules to be detected in the sample, to additionally move, e.g. on or in the cells, which could lead to longer imaging times (on account of relaxation processes of the cell dynamics) by at least the 10-fold, which could also induce a cell response, and thus to a falsification of the biological processes to be observed. Usually, stepper motors are used for this, which ensure a smoothened mode of movement by a rapid sequence of movement steps. "Constant" and "continuous" within the scope of the present invention means that there is no extended stopping of the sample during the measurement process (or a

measurement in the at-rest state, respectively), but
that the sample (or the sample holding means, respec-
tively) is always moved relative to the detection and
analysis system.

Preferably, the movement of the sample is con-
trolled directly by the detection and analysis system
in the x-y direction, it being possible to adapt such
controlling to the respective characteristics of the
detection and analysis system. If a CCD camera is used
in the detection and analysis system, the relative
shifting can be triggered directly by the frame shift
characteristic of the CCD camera. When a certain area
on the sample holding means is illuminated, which area
is being imaged on the entire pixel array used, the
sample is continuously shifted, and, simultaneously,
the image of the sample on the pixel array likewise is
shifted line by line by continuous frame shift. In case
of an optimum adaptation of the two speeds (relative
velocity of the movement of the sample and frame shift
(line readout speed) of the CCD camera), the informa-
tion collected by a labeled molecule of the sample
while traversing the illuminated region will be col-
lected by practically the same pixels. Optimally, the
speed with which the sample is moved will be equal to
the speed of the CCD camera, divided by the magnifica-
tion of the objective.

If, however, in addition to the x-y movement, also

the image along the z direction is sampled, preferably a separate control unit, in particular a unit having its separate pulse transmitter and its separate software, is used.

According to the invention, mainly fluorescence dye is used as dye, i.e. visualization is carried out by using epifluorescence microscopy. According to the present state, the best resolutions can be attained by this method; it is, however, also conceivable to carry out the method of the invention with other processes (e.g. RAMAN, infrared, luminescence and enhanced RAMAN spectroscopy as well as radioactivity), similar resolutions as those of fluorescence technology in principle being attainable with luminescence or enhanced RAMAN, yet above all with bioluminescence.

According to the invention, the use of the two-photon excitation fluorescence microscopy (Sanchez et al., J. Phys. Chem. 101 (38) (1997), pp. 7020-7023) has proven particularly suitable, since with this method it is also possible to efficiently circumvent the problem of the autofluorescence of many cells.

Furthermore, this allows for a practically background-free measurement, which can also speed up HTS analysis. The two-photon excitation fluorescence spectroscopy (or, generally, multi-photon excitation (Yu et al., Bioimaging 4 (1996), pp. 198-207)) is particularly suitable for a three-dimensional illustration of sam-

ples, resulting in a further advantage, above all with cellular systems.

In the embodiment with fluorescence spectroscopy, the arrangement according to the invention preferably comprises one or several of the following components:

· a laser as a precisely defined source of light, as well as

· acousto-optical switches with high specificity, by which the laser beam may rapidly (e.g., 10-20 nsec) be interrupted for a defined period of time,

· a processor which controls the switch, e.g. via a pulse program,

· a dichroitic mirror (which, e.g., reflects the exciting light upwardly towards the sample and passes the fluorescent light from the sample downwardly (towards the analysis system),

· a series of suitable filters known from conventional SDT arrangements,

· a mobile sample holding means, e.g. a processor-controlled x-y drive (stepper motor),

· a CCD camera by which the emitted light quantums which are passing the dichroitic mirror are converted into electrons and collected in pixels,

· a galvano-optic mirror which directs the image onto pre-selected (in x direction) adjacently arranged areas of the pixel array, perpendicular to the frame shift direction (y direction),

- a prism which divides the image into two spatially separated images with orthogonal polarization, and
- a processor which controls movement of the sample (of the sample holding means) by an x-y drive (stepper motor), by the signals from the CCD camera being used via an internal clock to trigger the movement.

According to the invention, it is also possible to stoichiometrically label different types of molecules with a dye, preferably a fluorescence dye, e.g. a receptor and a ligand, and to pursue both with the arrangement of the invention.

It is also possible to label at least two different types of molecules with different fluorescence dyes and to subject them to SDT analysis, wherein, in addition to the respective single fluorescence, also additional information can be obtained by determining, e.g., the Förster transfer (Mahajan et al., Nature Biotech. 16, (1998), pp. 547-552). However, it ought to be substantially emphasized that with the Förster transfer alone merely a (although highly selective) qualitative, yet not a quantitative information is possible, since this effect is highly dependent on the distance of the fluorophores (with $1/r^6$).

If cellular systems are to be assayed according to the invention, it is preferably started with cells of low autofluorescence, there being various cell types which have little autofluorescence from the begin-

ning (such as, e.g., mast cells or smooth muscle cells). Unfortunately, however, it is just the expression cells which, as a rule, are highly fluorescent, and therefore these or other cell types having intrinsic fluorescence must be provided in a low-fluorescent state by selected growing conditions or sample processing so that their autofluorescence will be brought to below a certain interfering level. When using two-photon excitation of fluorescence, this problem, however, does not occur from the very beginning, as has been mentioned before.

With the arrangement according to the invention, carrying out a visualizing method for single, e.g. biologically active, molecules is possible as a high throughput screening of biological units on the basis of the observation of single molecules (fluorophores).

High throughput screening (HTS) generally describes the search for certain "units" among a very large number of similar "units" (e.g. in a molecule library and a partial molecule library prepared by combinatorial chemistry). Such problems are encountered in many fields, both in basic bio-scientific research and also in the medically-pharmaceutically oriented industrial research and development. "Units", according to the invention, may be biological cells, yet also individual molecules or types of molecules, high throughput screening e.g. being possible for detecting rarely oc-

curring cells having a certain genetic defect. Besides
its usefulness in connection with questions of cellular
biology and pathology, high throughput screening is im-
portant in molecular biology. Thus, the arrangement ac-
cording to the invention may, e.g., be used to find
single DNA or c-DNA molecules in a sample comprising
many DNA molecules. In biochemistry, the separation of
macromolecules having certain properties, e.g. with re-
spect to ligand binding or state of phosphorylation in
or on cells, is a basic requirement which can be dealt
with according to the invention. The pharmaceutical in-
dustry needs high throughput screening both for select-
ing certain active agents and also for analyzing their
activity on biological cells. Each person skilled in
the art will know what belongs to HTS methods or which
materials can be used therefor (e.g. molecule libraries
prepared by combinatorial chemistry or genomic-combina-
torial libraries) (cf., e.g., "High Throughput Screen-
ing", John P. Devlin (Ed.) Marcel Dekker Inc. (1997)).

For a specific labeling of certain "units", ac-
cording to the invention mostly the natural principles
of the structurally-specific molecular recognition are
employed, such as the binding of antibodies or, gener-
ally, of ligands to receptor molecules. The preferred
use according to the invention of fluorescent ligands,
such as antibodies with bound fluorescence molecules,
allows for a both sensitive and selective detection of

units with receptors for the fluorescence-labeled ligands. As an alternative to fluorescent ligands, fluorescent groups can be inserted in protein sequences and coexpressed (e.g. the "green fluorescence protein" (GFP) or variants thereof ("blue fluorescence protein" – BFP).

According to the invention, with the use of fluorescence, a high throughput screening with simultaneous ultimative sensitivity (i.e. clear detection of the fluorescence of individual fluorescence markers) and high throughput rate (i.e., at least 10^6 (cellular) units per inch^2 per hour) can be realized. Chemical units (e.g. biological molecules, such as receptor agonists or antagonists) may be assayed without any problem with a throughput rate of at least 10^{10} or 10^{12} units per hour per inch^2.

When using cells in a HTS method, primarily microtiter plates are suitable with which a medicament screening can be carried out on complete cells, e.g. by titrating the cells into the individual wells which contain the substances to be screened (cf. e.g. WO 98/08092). Also the use or measurement of biochips (Nature Biotech. 16 (1998), 981-983) is possible with the system according to the invention.

If substances are identified as pharmaceutical target substances and isolated with the HTS method of the invention, which are new or for which so far a

pharmaceutical activity could not be demonstrated, the present invention, in a further aspect, relates to a method for preparing a pharmaceutical composition, which comprises mixing of the substance identified and isolated according to the invention with a pharmaceutically acceptable carrier.

According to the invention, a clear detection is considered to be given if the minimum signal/noise ratio determined for single molecules is more than 3, preferably between 10 and 40, in particular between 20 and 30. If the signal/noise ratio is below a value of approximately 2 to 3, interpretation of the information content of the measurement obtained may be a problem.

A specific variant of the method according to the invention is the combination with the flow cytometry technology, in which the cells are moved by a flow cytometer past the detection and analysis system. In the simplest instance, in a preferred variant of the arrangement of the invention, a flowthrough cell is provided with the sample holding means (or as the sample holding means itself, respectively).

As has already been mentioned, the arrangement according to the invention is particularly suitable for the analysis of samples which comprise biological cells, wherein particularly HTS methods may be carried out efficiently with the arrangement according to the invention. The spectrum of use of the arrangement of

the invention is, however, also highly efficiently applicable to cell-free systems.

In the arrangement according to the invention, the relative shifting between sample and the highly sensitive (high- resolution) detection and analysis system preferably is controlled by the detection and analysis system itself, in particular by the CCD camera, if such relative shifting is to take place continuously, which is advantageous particularly in case of a lateral scan.

Since fluorescence analysis at present yields the best analyses, the arrangement according to the invention preferably comprises an EPI fluorescence microscope. Moreover, control of the continuous relative shifting can be triggered via the frame shift of the CCD camera, control being directly effected through the CCD camera, or in parallel by a synchronisation mechanism (e.g. location-correlated via photodiode triggering signals by using a co-transported punched tape, such as, e.g., described in Meyer et al., Biophys. J. 54 (1988), pp. 983-993).

A preferred embodiment of the present invention therefore is characterized in that the sample movement and the frame shift of the CCD camera are synchronized with each other by location-correlated signals derived from the continuous sample movement, preferably by using a punched tape moved together with the sample, and a fixed photodiode which transmits a signal when pass-

ing a punched hole.

In a further aspect, the present invention relates to a method for visualizing molecules, interactions between molecules, and molecular processes in a sample by using the SDT method employing an arrangement according to the invention.

Therefore, the present invention also relates to a method for visualizing molecules, their movement, molecule interactions, and molecular processes in a sample, wherein a sample in which certain molecules have been labeled with marker molecules are introduced into an arrangement according to the invention, the sample is imaged on a pixel array by the CCD camera, the sample and/or the detection and analysis system being shifted relative to each other by utilizing the frame shift of the CCD camera so that the signals of each single molecule in the sample will be collected in the same pixels after having been converted into electrons, until the single molecule signal exceeds a certain minimum signal/noise ratio.

Preferably, the relative movement of the sample is directly controlled according to the frame shift of the CCD camera, the relative movement of the sample being effected in lateral direction, preferably constantly and continuously.

In a further aspect, the present invention relates to a method for quasi-simultaneous imaging of fluores-

cence-labeled molecules in their distribution over complete biological cells (or biological systems, respectively) and for pursuing molecular movements and processes by repeating this imaging at temporal intervals by using the SDT method which is characterized in that a sample with cells, in which certain molecules have been labeled with marker molecules, are introduced into an arrangement according to the invention, the fluorescence image for a focussing plane is imaged on the pixel array of the CCD camera, the focussing plane is shifted step-wise along the z direction by a piezo-element, the fluorescence images to each plane being separately arranged on the pixel array, and after imaging of all the focussing planes, the image of the fluorescence labeled molecules in the cells is calculated, whereupon optionally the images of the focussing planes are repeated so as to illustrate molecular movements and processes by serially arranging images of all the focussing planes.

    With this method, not only detection of single molecules on cell surfaces or in cells can be effected with the arrangement of the invention, but it is also possible to pursue the processes in (live) cells down to molecular movements and processes in terms of space and time. Thus it has become possible for the first time to image live cells in "real time" and thus observe molecular processes in and on these cells.

Of course, this method is not only usable for complete cells, but also for observing processes in all biological systems, such as, e.g., in isolated cell membranes or in synthetic cell compartments or synthetic membranes in which biological molecules are incorporated (according to the invention, all these systems are also encompassed by the term "biological cells").

Preferably, imaging on the pixel array of the CCD camera, primarily in a 3D scan of the cells, is effected at a rate of from 1 to 3 ms per image and at a capacity of up to 300 images per array, with an image size of 80 x 80 pixels. Other adjustments can be further optimized by the skilled artisan for the respective CCD camera, source of light etc. used, in dependence on these individual components.

With the arrangements according to the invention and by means of the methods of the invention it is not only possible to use a single fluorescence marker, but the use of two or more fluorescence markers is possible without any problem. For instance, also the system described in US 5,815,262 in principle can be employed according to the invention.

According to a preferred embodiment, the present invention also relates to a method in which at least two different types of molecules in the sample, in particular in the cell, are labeled by at least two dif-

ferent fluorescence markers, whereupon not only the movement of one molecule in the system, but also the relative movement of the different molecules in the system can be imaged and pursued in terms of time and space.

Preferably, the fluorescence image is captured for two orthogonal polarization directions for each fluorescence marker by dividing the image into two images with orthogonal polarization direction. This may be enabled by using a Wollaston prism and an imaging optic which has a parallel beam region, the Wollaston prism being used in the parallel beam region of the source of light.

In addition, also a galvano-optical rotating mirror may be used in the parallel beam region, e.g. of an epifluorescence microscope.

By using the rotating mirror and the Wollaston prism, in a 3D scan successive images of the focussing planes with both polarization parts can be stored separately adjacently on the entire width of the pixel array of the CCD camera. By means of frame shift, this image sequence can be shifted as a whole by one image width, whereupon the next image sequence will be stored by mirror rotation until either sufficient information has been gathered or the pixel array is full. Then the entire information can be read out for processing to a 3D image, and the camera will be free for the next 3D

imaging.

In doing so, positioning and shifting of the im-
ages to each sample position on the pixel array of the
CCD camera for different fluorescence phases and two
polarization directions can be effected by the control
unit by means of a pulse transmitter and corresponding
software.

Preferably, cells of low intrinsic fluorescence
are used in the sample.

Preferably, the method according to the invention
is carried out as a high throughput analysis, wherein,
e.g., a molecule library can be analyzed as the sample,
preferably a molecule library prepared by combinatorial
chemistry. According to the invention, also the inter-
action of an entire molecule library with biological
cells can be analyzed.

The fields of application for the present inven-
tion are practically unlimited, preferred are, however,
pharmacy (primarily HTS of new chemical units) as well
as biochemical questions, since, due to the extremely
high sensitivity of the methodology according to the
invention (a single molecule can be pursued) and the
exact localization (e.g. with a precision to at least
30 nm) basically each individual molecule or molecule
associate, e.g. on or in cells, can be detected and
identified (optionally also isolated). Thus, the bind-
ings of all natural ligands to a cell (hormones, pri-

mary messenger substances, etc.) or cell-cell recogni-
tion molecules with molar binding can be analyzed, also
as regards the exact binding kinetics and binding con-
formation, as well as regards the mobility of these
components within the cell or within the cell membrane
(analogous to Schmidt et al., J. Phys. Chem. 99 (1995),
pp. 17662-17668 (for molecule position and mobility de-
terminations); Schütz et al., Biophys. J. 73 (1997),
pp. 1-8; Schmidt et al., Anal. Chem. 68 (1996), pp.
4397-4401 (for stoichiometric determinations); Schütz
et al., Optics Lett. 22 (9), pp. 651-653 (as regards
conformation changes).

Furthermore, the system according to the invention
is particularly suitable for analyzing and identifying
or isolating, respectively, (alternative) binding part-
ners in receptor-ligand or virus-receptor systems,
wherein also potential agonists/antagonists and their
action (e.g. the competitive inhibition) can be pre-
cisely analyzed. This is particularly essential when
finding new chemical units (NCU) in the field of me-
dicament screenings.

When analyzing entire cells, the focus plane may
be varied; in a rapid variant, a section through the
cell (preferably, the upper cell half; "lower" meaning
the side facing the sample holding means) is analyzed.
Thus, it is also possible to analyze complex processes
in a cell, such as nucleopore-transport, the effect of

pharmaceuticals with a target in the cell or secondary reactions in the cell, on single molecule level.

According to a preferred embodiment, the system of the invention may also be used to analyze three-dimensionally (3D) occurring processes in single cells, such as cells which have been pre-selected in a first area scan according to the invention. In doing so, by a continuous or discrete shift of the focus plane along the z axis, in addition to the inventive mode of procedure (sample shifting with synchronized frame shift of the CCD camera), the three-dimensional arrangement of fluorescence-labeled molecules or associates on or in the cell can be imaged, in measurement times in the range of seconds or even therebelow, with a location resolution close to the diffraction limit. Compared to the hitherto only other method, the confocal scanning fluorescence microscopy, CSFM (Handbook of Biological Confocal Microscopy, ed. James B. Pawley, second edition (1995), Plenum Press, New York and London), the illustrated, above-indicated method according to the invention, firstly, is more rapid by at least a factor 1000, since simultaneously the information with equal resolution can be collected by at least 1000 focus areas, whereby, secondly, it is possible for the first time to image non-static molecules or associates, respectively, in spatial-temporal arrangement in periods of time (1 s, e.g.,) which are small enough to observe diffu-

- 27 -

sion processes, energy-driven movements or metabolic processes.

In a preferred embodiment, thus, the focus plane of the detection and analysis system (in particular, of the epifluorescence microscope) can be shifted along the z direction (i.e., normal to the x-y plane which is defined by the sample surface (the sample holding means)), optionally in addition to the relative movement between sample and detection and analysis system.

In doing so, 3D imaging is carried out, preferably by imaging of discrete, consecutive focus planes in z direction, in rapid cyclical repetition, during a continuous relative movement between sample and CCD camera, by parallel collecting of the images of different z planes on the pixel array by using a galvano-optical mirror. Thus, substantial advantages of both imaging methods can be combined, which preferably is used for cellular HTS, yet also in general for molecular-mechanistic questions of cellular biology, physiology and pharmacology.

Preferably, the x-y scan and the 3D imaging can be effected simultaneously. To this end, the images of each z plane can be captured adjacently by using the galvano-optical rotating mirror. In the slow x-y scan, several z cycles are passed per illumination time of each fluorophore. By this combination, the x-y scan is slowed down, i.e. by the factor of the number of the z

- 28 -

planes.

The method according to the invention and the arrangement according to the invention are also very suitable for detecting the specific binding of labeled nucleic acids on so-called arrays. In doing so, a plurality of different nucleic acids (e.g., cDNAs, ESTs, genomic sections with various mutations (SNPs)) are immobilized in uniform patterns on a surface (e.g. synthetic material or glass). These arrays are then incubated with the sample to be tested comprising fluorescence-labeled nucleic acid molecules, the molecules from the sample being capable of specifically hybridizing with their homologous counterparts. This can be repeated with various markers on the same sample. In the prior art, evaluation of the binding events hitherto frequently has been effected with scanners or imaging methods which have a relatively low resolution and sensitivity. Here, the system according to the invention offers clear advantages, since the enormous speed as well as the high spatial resolution of the SDT analysis according to the invention come as an addition to the ultimative sensitivity of the method of the invention. Thus, it is easily possible to adapt systems as described in WO 97/43611, e.g., with the system according to the invention and to analyze them according to the invention.

This is primarily advantageous if the concentra-

tion of the labeled nucleic acids of the sample is very
low. Thus, e.g., mRNAs which are present in the cell in
a very low copy number (low abundance mRNAs), reliably
can be detected with a suitable array. Further applica-
tions of this specific aspect of the present invention
relate to problems in which the amount of the nucleic
acids of the sample is very low, such as in forensic
trace analysis, or in an analysis of embryonic or stem
cells.

Moreover, the method according to the invention is
particularly suitable for detecting nucleic acids in
the so-called in situ hybridization. In this instance,
tissue slices are incubated with a labeled sample. The
specific binding of these nucleic acids of the sample
allows for a statement as to which mRNAs are expressed
in which regions of the tissue section. Since these
mRNAs to be detected often are present in very low copy
numbers, a high sensitivity of the detection system as
is provided by the method according to the invention is
advantageous.

Analogous to in situ hybridization with nucleic
acids, also biorecognitive molecules, such as antibod-
ies, can be used as sample molecules, in which instance
the epitopes (e.g., certain protein molecules) recog-
nized by the antibodies can be detected with high sen-
sitivity.

Likewise, the method of the invention can be used

in the analysis of chromosomes. In doing so, chromosome preparations are prepared on a carrier, and these are incubated with a corresponding nucleic acid sample. Detection of a specific binding allows for a conclusion regarding the localization of individual genes on the chromosomes.

The invention will now be explained in more detail by way of the following Examples and drawing figures, without, however, being restricted thereto.

Fig. 1 shows the usual configurations of units for high throughput screening;

Fig. 2 shows one possible arrangement according to the invention;

Fig. 3 shows the relative movement of the sample with frame shift;

Fig. 4 shows the screening of units on surfaces or in multi-well plates;

Fig. 5 shows the screening in a laminar flow cell;

Fig. 6 shows the relation between screening time and resolution;

Fig. 7 shows the analysis of detected units;

Fig. 8 shows the positions of labeled molecules and the temporal tracing of molecule positions;

Fig. 9 shows the molecular association, co-localization, stoichiometry from signal quantization;

Fig. 10 shows the conformation change on the single molecule;

Fig. 11 shows the ligand binding;

Figs. 12-13 show the co-localization of two dif-
ferently labeled ligands by energy transfer (Fig. 12),
or by comparing the positions of the two dye molecules
(Fig. 13);

Figs. 14-18 show the detection of individual lipid
molecules in native cells;

Fig. 19 shows the microscopy of individual lipid
molecules with two-photon fluorescence excitation;

Figs. 20A, B and C show the three-dimensional im-
aging of a selected single cell with single fluorophore
resolution;

Fig. 21 shows an arrangement according to the in-
vention, suitable for 3D-analysis of complete cells;

Fig. 22 shows the reading-in of the images in the
pixel array of the CCD camera;

Fig. 23 shows the operating mode in large-area
screening.

E x a m p l e s :

**E x a m p l e 1: Arrangement according to the in-
vention, employing fluorescence microscopy**

Conventionally used configurations of units for
high throughput screening (HTS) are illustrated in Fig.
1, which are all employed as measurement arrangements
in the method according to the invention. Usual mole-

cule libraries, prepared by combinatorial chemistry,
are assembled on small (0.2 to 0.4 mm) polymer beads
each carrying a single molecule species (cf., e.g.,
Devlin (1997), pp. 147-274). The measurement arrange-
ment according to the embodiment described above starts
from a conventional fluorescence microscope (Fig. 2) by
means of which fluorophores (8) present on the sub-

strate surface (7) in the illuminated area (~100μm^2)
could be individually detected and their movement could
be followed, with a signal-to-noise ratio of ~30 for
single fluorophores (published in Proc. Natl. Acad.
Sci., USA (1966) 93: 2926-2929). A Zeiss microscope
(Axiovert 135-TV) having a x100 objective (5) (Neo-
fluar; numeric aperture = 1.3, Zeiss) was used. For the
fluorescence excitation, the laser light of the 514 nm
line of an argon$^+$ laser (1) (Innova 306, coherent),
which was operated in TEM$_{00}$ mode, was coupled through an
acousto-optical modulator (1205C-1; Isomet) in the epi-
port of the microscope. A λ/4 plate delivered circular-
polarized excitation light. By using a defocussing lens
(3) (f = 100 mm) in front of the dichroic mirror
(515DRLEXT02; Omega), the Gaussian excitation profile
was set to 6.1 $\pm$ 0.8 μm full width at half maximum

(FWHM) and 57 $\pm$ 15 kW/cm^2 mean excitation intensity.
The illumination time for each pixel array image was
5 ms. After long-pass filtering (10) (570DF70 Omega and

- 33 -

OG550-3 Schott), the fluorescence was imaged by a lens
(13) onto a liquid-nitrogen-cooled CCD camera (15)
(AT200, 4 counts/pixel read-out noise; Photometrix),
equipped with a TH512B chip (14) (512 x 512 pixel,
27 μm^2 pixel size; Tektronix). The point-transfer func-
tion of the microscope was described by a two-dimen-
sional Gaussian intensity distribution with a width of
0.42 μm FWHM, as was found by determining images of
30 nm fluorescent beads (Molecular Probes). The dif-
fraction-limited area thus was 0.14 μm^2. With 0.48 ±
0.08 μm FWHM, the width of intensity profiles for sin-
gle molecules was larger than the point-transfer func-
tion of the microscope, the additional broadening
having been caused by molecular diffusion. The CCD was
used as a memory means, with 12 consecutive images of
40 x 40 pixels being captured, wherein up to 140 pixel
arrays could be imaged per second, due to CCD frame
shift. This frame shift is used according to the inven-
tion for continuous movement of a sample holding means.

   According to the present invention, this measure-
ment principle can be applied to biological samples
with fluorescent ligands in configurations as illus-
trated in Fig. 1. According to the apparatus by which
the invention has been realized, sample screening is
enabled which has a constant single fluorophore sensi-
tivity with maximized throughput rate. The basic idea

is once more explained in Fig. 3. At constant illumina-
tion of an area which is imaged on the total pixel ar-
ray used, the sample is continuously shifted and,
simultaneously, the image of the sample on the pixel
array by continuous frame shift, line per line. With as
precise a coordination of the two velocities (v (sam-
ple) = v(CCD)/magnification of the objective) as possi-
ble, the fluorescence collected from a fluorophore of
the sample will be collected during traverse of the il-
luminated area by practically the same pixels.

In Figs. 4 and 5, the cumulating image of a
fluorophore up to reaching the read-out side of the
pixel array is outlined for screening configurations
according to Fig. 1. Optimization of the numerous appa-
ratus variables and parameters is possible for the
skilled artisan in analogy to the known methods; the
ratio between resolution and measuring time is shown in
Fig. 6: For typical characteristics of obtainable CCD
cameras, sources of light and objectives, the measuring
time for the screening of an area of 1 inch^2 was calcu-
lated, as a function of the resolution. Basically,
there is a clear-cut region of the optimum relationship
between measuring time and resolution, which in the ex-
ample chosen is in the range of measuring times of from
15 to 30 min for 1 inch^2 (6.45 cm^2) sample area, at a
resolution of from 1-0.5 μm. The working point on this
curve is adjusted by binning. To this end, the informa-

tion of neighbouring pixels is combined (e.g. of b x b
pixels), whereby the resolution decreases, at increas-
ing maximum velocity v (frame shift) of the CCD camera
and slightly increasing sensitivity. The latter is
based on the fact that the noise merely is due to read-
out noise, and thus is equal in amount for the read-out
of the counts in a single pixel as for the read-out of
the counts in b x b pixels. The imaging quality of in-
dividual fluorophores thus is substantially maintained
during screening (in the example according to Fig. 6,
250 counts per fluorophore are collected at 5 counts of
read-out noise).

By the continuous sample movement, the waiting
time is minimized which is necessary at discontinuous
sample movement, due to the movement forming in the
sample when the velocity is changed. Merely after ter-
mination of a line scan, the sample has to be returned
and shifted by the width of the illuminated area, so as
to collect the next line scan (in the example of
Fig. 6, a waiting time of 1 s was allowed therefor).

The inventive combination of ultimative sensitiv-
ity and comparatively very rapid sample throughput
opens up new fields of application. With a screening
time of ~15 min of a sample of typical size, a time
range has been attained which allows for a screening
under generally constant conditions of the samples.
Samples with a correspondingly short life time can be

assayed, and the detected units can be further used or
analyzed. This, moreover, allows for the use of a wide
range of fluorescence ligands with appropriately rapid
dissociation rates (e.g., weakly binding antibodies).
The simultaneous single fluorophore sensitivity basi-
cally enlarges the field of application to situations
in which labeled sites per unit sought are to be ex-
pected in low numbers (down to a single site, such as
when finding a mutation in a DNA sample).

According to the invention, the rapid and sensi-
tive screening described can be further combined with a
high selectivity and specificity. To this end, selec-
tive excitation of the fluorescence markers by two-pho-
ton absorption is used, whereby the fluorescence
collected almost entirely comes merely from the thus-
excited fluorescence markers in the focus area. In a
further mode of action, two fluorescence-labeled
ligands are used simultaneously, which both have neigh-
bouring binding sites on the target structure. This
may, e.g., be a natural ligand of a receptor merely oc-
curring in the unit sought, together with an antibody
which binds to the receptor molecule in the vicinity of
the ligand. As outlined in Fig. 12, in case of a selec-
tive excitation of one of the two fluorophores (donor)
and collection of the fluorescence (by appropriate op-
tical filters) of only the second fluorophore (accep-
tor), merely the fluorescence formed by the transfer of

- 37 -

energy from the donor to the acceptor will be detected.
For this purpose, both fluorophores have to be in the
immediate vicinity (distance ≤ 8 nm). In this way,
ligand pairs specifically bound to receptors become de-
tectable individually and highly selectively (with
still a high signal/noise ratio). Alternatively, sepa-
rate images of both dyes can be made (with a time lag
of merely 5 ms, cf. Fig. 13). Co-localized dye mole-
cules (within the position precision of ~50 nm) allow
for a highly specific allocation, since here the quan-
tum information of the single molecule intensities of
two different fluorophores are retained as criterion,
in contrast to the transfer of energy. Besides increas-
ing the selectivity of fluorescence by transfer of en-
ergy, the specificity of the signal can be increased by
illumination in total reflection (cf. Fig. 3 below).
Thus, only those fluorophores are excited which are lo-
cated in a range of approximately 100 nm from the sub-
strate surface (exponentially fading light intensity).
The detection sensitivity also reaches single fluoro-
phores. This type of illumination shall complement the
invention by enabling the use of high throughput
screening for units (mainly cells) having a high auto-
fluorescence.

    According to the invention, immediately after the
detection of sought units by screening, the same appa-
ratus allows for a detailed analysis of these units.

This may either be carried out directly in the screen-
ing sample, or after transfer of a unit into an analy-
sis cell. Fig. 7 shows this on the example of a
biological cell.

The analysis cell allows for single molecule mi-
croscopy in a region of the biological cell which is
freely accessible to an exchangeable buffer solution
and active substance. Furthermore, the cell is practi-
cally tightly bound electrically to the substrate so
that the highly sensitive fluorescence microscopy can
be combined with electrophysiology, e.g. for observing
single ion channels, electrically and optically.

Figs. 8 - 13 outline five basic types of informa-
tion which become possible by single molecule micros-
copy on transferred units. In this connection, binning
is employed to adapt the temporal and lateral resolu-
tion to the desired information. The sample is not
moved, and short (ms), periodically repeated illumina-
tion is employed. This allows for each illumination to
detect the positions of sufficiently far removed single
fluorophores and to follow them temporally (Fig. 8).
Thus it can be decided whether a labeled receptor is
mobile, restricted mobile, or immobile, diffuses freely
or has limited diffusion or is self-associated, co-as-
sociated with other components, or transiently clus-
tered. Also the distribution over the (cell) surface
can be made visible. The high signal and the high sig-

nal/noise ratio S/N (approximately 150 counts and S/N = 30 for 5 ms of illumination) allows for the allocation of observed signals to the number of co-localized fluorophores. This opens the field for numerous mechanistic studies relating to the association, co-localization and stoichiometry of associated components, outlined in Fig. 9 for dimerization of a membrane component.

Special ligands (whose fluorophore points into a fixed direction after binding to the receptor) can be employed for single-molecular detection of conformational changes. A slight rotation of the fluorophore with a structural change of the receptor suffices to detect the conformational change via the intensity change of its fluorescence signal, as is outlined in Fig. 10. For this, both linearly polarized light of different directions of polarization and circularly polarized light are used.

For ligand concentrations of a few nM at the most, ligand binding can be analyzed on single-molecular level (Fig. 11), including the stoichiometry of the ligand binding, as well as allosteric and cooperative effects at ligand binding. By using two different fluorescence-labeled ligands, highly specific statements can be made with a high reliability that the ligands observed are bound to the receptor. For this, either energy transfer between the two fluorophores can be

used (Fig. 12), or their co-localization in consecutive images for each of the two dyes. (Fig. 13).

The inventive continuous imaging of the fluorophores in the sample by synchronous movement of the sample and CCD frame shift (according to Figs. 3 to 6) has neither been described nor suggested in the prior art relating to single fluorophore imaging, since there, only static images have been captured in immobile samples. With the system according to the invention, in addition to the ultimative optical resolution and sensitivity of the time-resolved detection of single molecules (e.g., receptors on cells), a considerable screening speed has become possible which is at least 1000 times more rapid than in the alternative methods of confocal microscopy, with simultaneous observation of an ensemble of molecules which is not possible by confocal microscopy.

E x a m p l e   2 : Detection of fluorescence-labeled lipid molecules in the plasma membrane of native smooth muscle cells

Methodology: smooth muscle cell, HASM: human aorta smooth muscle, stable cell line of wild type, are allowed to grow on a cover glass and subjected to microscopy in PBS buffer. Incorporation of DMPE-Cy5 (dimiristoyl-phosphatidyl-ethanolamine with Cy5 (from AMERSHAM) bound as dye molecule) is effected via lipid vesicle (POPC: palmitoyl-oleoyl-phosphatidylcholine,

from AVANTI). Each 1000th lipid in the vesicles was a
DMPE–Cy5 (mean: 5 DMPE–Cy5 per vesicle). Addition of
these vesicles via the flowthrough cell to the HASM

cells in the microscope (50 µg/ml vesicle, incubation
for 10 min, then washed out with PBS buffer) leads to
DMPE–Cy5 individually incorporated in the plasma mem-
brane, via vesicle/cell membrane/lipid exchange. This
process of the delivery of one DMPE–Cy5 to the plasma
membrane is directly visible in Fig. 16, the vesicle
(at ~ 10 DMPE–Cy5, cf. high signal) quickly diffusing
along the cell membrane and one DMPE–Cy5 suddenly
changing over from the vesicle into the plasma mem-
brane. The movements of the lipid sample and of the
vesicle could be observed separately (cf. trajectories
in Fig. 16, bottom). Such an exchange could not be ob-
served previously on single molecule level. What is es-
sential, however, is that the intensity of one
fluorophore is still clearly resolved in the cell hav-
ing autofluorescence (Fig. 14). In the present example
(Figs. 14–18), the intensity of the laser light (630
nm) was reduced such that the effective fluorescence
background of the cell became less than the readout
noise. The intensity may, however, be increased at any
time, so that – via the intrinsic fluorescence of the
cell – it is possible to get an orientation regarding
the site at which the measurement is being carried out.
The intensity distribution of 300 single molecules

(Fig. 15) resulted in a signal/noise ratio of 25 for
the detection of individual molecules in the native
cell membrane, at 5 ms of illumination. For a better
understanding of the peaks shown: The area shown com-
prises 576 pixels, and this corresponds to an object
area of ~6 x 6 μm (each pixel is 27 x 27 μm, a x 100
objective was used). The HASM cell has a length of ap-
proximately 100 μm, a width of 15-20 μm, and a height
of 5-10 μm. The illustration is diffraction-limited,
i.e. each dot source is imaged as a Gaussian spot hav-
ing a radius of ~270 nm, this corresponds to 1 pixel
(60% of the peaks on 4 pixels). In the peak there were
152 cnts.

As a matter of routine, sequences of up to 14 im-
ages were captured (cf. Fig. 16), 5 ms illumination
each, with dark intervals of between 10 to 30 ms (i.e.
measuring times of up to ~0.5 sec). These result in
trajectories for the movement of the labeled lipids in
the plasma membrane. ~300 of such trajectories were
evaluated (includes measurements on three different
cells and on different locations of the cells, yet al-
ways on the upper side of the cells which are adhered
on the bottom of the cover glass). In the measuring
time of 0.5 sec, no convection or other cell movement
was observed (except for a few erratic cell jerks). The
result was impressive: The evaluation of the trajecto-

ries is illustrated in Fig. 17: The square of the dis-
tance (MSD = mean square displacement) between observed
molecule positions and trajectories is entered against
the respective time interval Δt. With Brown's diffusion
processes, this should result in a linear connection;
MSD = $4D_{lat} \cdot \Delta t$, with the diffusion constant D_{lat} for
lateral movement resulting from the ascent $4D_{lat}$. At
first, for short diffusion lengths, a diffusion with
D_{lat} = 0.6 μm^2/sec appears, a typical value for lipid
diffusion in cell membranes (from ensemble measurements
via FRAP : fluorescence recovery after photobleaching).
For longer diffusion periods, the range of movement of
the lipid sample remains limited. The dashed line indi-
cates that the sample in its movement remains restriced
to an area having a radius of 300 nm. This is the first
direct proof of the existence of lipid domains which
hitherto have merely been postulated ("lipid rafts",
Simons and Ikonen, Nature 387: 569-572). For "lipid
rafts", the preferential partition of lipids with satu-
rated acyl chains is postulated (as in the lipid sample
DMPE-Cy5). In fact, the homologous sample DOPE-Cy5
which has only one double bond in each acyl chain, did
not exhibit a partition in domains (proven by co-local-
ization, not illustrated), but free, unrestricted dif-
fusion in the plasma membrane (Fig. 18).

A further analysis of the domains showed that they

are anchored on the cytoskeleton and move actively (uni-directional).

In principle, these results indicate that any application of SDT on cells, also as hitherto has been used on model systems, opens up new possibilities, simply due to the fact that processes can be viewed on a single-molecular basis and dynamically, which so far have been accessible merely via ensemble-mediated data. In this connection, the present proof is essential that single fluorophores on live cells (at least on these smooth muscle cells) can be viewed by microscopy clearly and with time resolution. The marker, Cy-5, may also be attached to a ligand having the same fluorescence properties. The frame-sample-shift reduces this resolution only unsubstantially, it serves for the continous screening of complete cell cultures or cells in nanotiter plates etc. Resolution can be further improved by two-photon-excitation fluorescence microscopy. Fig. 19 shows the first realization of a two-photon imaging of two phospholipids (PE-) with bound TMR (tetramethyl-rhodamine) as fluorescence markers in a phospholipid (POPC) membrane.

E x a m p l e 3 : Simulation of a real time image of the distribution of single fluorophores on complete cells, the spatial-temporal resolution, the position precision and the detection safety of the fluorophores

In Example 2, the observation of the single lipid

diffusion in the plasma membrane of the cell i.a. was possible because the plane of the lipid movement (membrane surface) could be brought into register with the focus plane (layer with an effective thickness of

1.6 μm) to a sufficient extent, which was realized by focussing on the upper rim of the cell.

To capture movements in any direction, also transverse to the focus plane, at any location on or in the cell, as well as almost at the same time for all the fluorescence-labeled molecules, the sample movement according to the invention and frame shift is carried out in the following variant:

Methodology: The methodology and the arrangement for imaging is as in Example 2, with two substantial differences:

1.) For sequential imaging, the focus plane is moved through the cell step-wise from position to position in the z-direction. This is outlined in Fig. 20A, with the focus layer in red, (effective thickness from which fluorescence photons are collected is 1.6 μm), a cell in green (approximate height of 8 μm), with an ensemble of randomly distributed equal fluorophores (black dots).

2.) A CCD camera of the following specifications is being used: with large, elongate pixel array and particularly rapid frame shift (e.g. a CCD camera with

2048 x 256 pixels of a size of 20 x 20 μm, 7 μs shift

time per line, and 2 μs/pixel readout time, conversion

of 0.8 electrons/photon, as is offered for spectroscopy

by PHOTOMETRICS, e.g.).

In the exemplary embodiment (Fig. 20), 20 images
are captured while the focus layer passes through the
entire cell. Each image is taken with the same illumi-
nation time "t(ill)" on the same partial area of the
pixel array (gray area in Fig. 20A with 100x256 pix-
els), and then shifted by frame shift, as illus-
trated in Fig. 20A for the first, second and last
image. During the time required for the frame shift
("t(fs)", 0.7 ms with the above-specified camera, imag-
ing is interrupted (by interrupting the illumination or
covering the camera).

Shifting of the focus plane per image captured is

chosen to be equal to 0.4 μm in the exemplary embodi-

ment, so that the 20 images will just cover the entire

cell of 8 μm height (cf. Fig. 20A). The time "t(ill)"
is freely selectable within limits. On the one hand,
"t(ill)" should be substantially longer than "t(fs)" to
keep low the information loss due to the illumination
pauses. On the other hand, the entire imaging time
t(total)=(t(ill)+t(fs))x 20 should not be too long
(long "t(ill)"-times are advantageous to even out un-
specific fluorescence) so that the molecule ensemble of

– 47 –

the entire cell can nearly be imaged in real time. Real time imaging requires t(total)<t(mov), wherein "t(mov)" is the time which is required by the molecules imaged to move over a longitudinal extension which corresponds to the optical resolution (approximately 0.5 μm in x- and y-direction, and approximately 1.6 μm in z-direction). For diffusion of typical membrane proteins or of actively transported components, "t(mov)" is mostly below approximately 0.6 s. From this there results for the exemplary embodiment a range of 5 ms < t(ill) < 30 ms for real time imaging of the fluorophores of a cell. The entire imaging will then take 0.1 s < t(total) < 0.6 s.

Fig. 20B illustrates the calculated result of real time imaging of fluorophores on a cell membrane for t(ill)=5 ms, wherein the data and conditions measured in Example 2 were taken as a basis for single images (3 counts/pixel of autofluorescence of the cell, 152 counts/fluorophore for 5 ms illumination with an intensity averaged over the focus depth, lateral resolution of 0.5 μm, and focus depth of 1.6 μm). These data set all the parameters for calculating the 3D image for the above-described method, a rotation ellipsoid with random deviations being chosen as the cell form (Fig. 20A shows the front view of the cell), and with randomly distributed fluorophores on the cell mem-

brane (black dots, total of 45). Figs. 20B and C show the front view of the 3D fluorescence image of the cell and of the fluorophores, produced by complete simulation of the imaging method (in the simulation, each fluorophore emits fluorescence photons corresponding to the illumination at the moment, which are imaged in random distribution on corresponding pixels, taking into consideration the diffraction-limited imaging function of the microscope, after corresponding conversion to electrons as "counts"). The color code chosen shows green for a low count number (autofluorescence of the cell and readout noise), and yellow, light red to dark red for increasing number of counts. The color code is chosen such that the light red range approximately reproduces the resolution volume (due to the collection statistics slightly inexact ellipsoid with diameters of approximately 0.5 μm in the x-y plane and

1.6 μm in z-direction), and that the dark red core reflects the precision of the positioning of individual dyes (approximately 50 nm in the x-y plane and 150 nm in z-direction). The loss of data on account of dark periods during the frame shift could be approximately taken into consideration by interpolation between the intensities of consecutive images by the fact that in the mean, each fluorophore appears in 4 images (focus layer thickness = 1.6 μm = 4 x dz, with dz = image-to-

- 49 -

image shift = 0.4 μm).

The simulation shows that 3D cell images in real time and with a clear detection of the fluorescence-labeled molecules are possible with the method according to the invention. In principle, this 3D image of the cell can be repeated several times after a minimum time each ("t(read)" which is necessary to read out the pixel arrays (approximately 1 s with the above-mentioned camera)). The number of repetitions is limited by bleaching of the fluorophores. For the conditions assumed in Example 2, at least a 3-fold repetition is possible without large losses by bleachout. In principle, the method cannot only image fluorophores on cells, as is illustrated in the exemplary embodiment, but also fluorophores in cells. To this end, the use of two-photon excitation may be advantageous, at least for studies in cells with high or not minimized autofluorescence.

Such real time 3D images of molecule ensembles of a cell do not only go far beyond the prior art, but in terms of quality they open up new ways for analysis of cyto-physiological processes, such as the uncovering and analysis of component organization/reorganization as an essential basis for the spatial-temporal regulation and coordination of cellular processes, or the mechanistic analysis of morphological responses of the cell to an external stimulus by, e.g., a messenger sub-

stance or by a pharmacologically active substance, or by a possible active substance identified according to the invention according to Example 1.

E x a m p l e 4 : Arrangement for 3D imaging of complete cells

Fig. 21 is a representation of the SDT microscope according to the invention, with which in particular a 3D scan of complete biological cells may efficiently be carried out.

The source of light here consists of one or several lasers (argon laser, dye laser, two-photon excitation laser). The AOM (2) (acousto-optical modulator) allows the laser light to pass for adjustable times (controlled by the controller (19)) and certain colors, such as merely the light of the argon laser for certain exposure time, repeatedly after adjustable dark pauses, or alternatingly, laser light of different wave lengths (argon laser line, e.g., 514 nm, and 635 nm from the dye laser, which is pumped through the argon laser) to excite two different dye molecules in the sample.

The lens (3) widens the focus plane (9); it is exchangeable for illuminating 20 μm to 600 μm areas. For large exposure surfaces (in SDT-x,y-scan mode), the light is guided through a single mode-fiber optic bundle between AOM (2) and lens (3), so that homogenous illumination of round or rectangular regions (SDT-x,y-scan) is attained (not included in the Figure). The fo-

cus plane (9) is adjusted to a certain z-value (distance between focus plane and sample carrier surface (7)) by means of a z-Piezo (18) which shifts the objective (5) in z-direction.

The sample is horizontally shiftable (in x,y-plane) by precision motors (17) which shift the sample stage (6) and are controlled by the controller (19). In this connection, the z-piezo (18) may be used for secondary regulation of a constant z-value (via capacitative distance measurement, not included in the Figure).

The fluorescence of individual dye molecules may be detected on the pixel array (14) of the CCD camera (15) as a localized and clearly resolvable signal. At first, the fluorescence of a molecule is refracted by the objective into parallel beams (in the meantime state of the art for objectives of all fluorescence microscopes) of a certain angle. The bundle of beams passes the dichroitic mirror (4) which merely reflects the excitation light, as well as filters (10) which are merely transparent for the fluorscence light.

The next element in the beam path is a galvanometer-mirror (11). The mirror can be adjusted to any angle (necessary region approximately +/- 5 degrees, corresponds to the width of the CCD camera indicated below by way of example) by the controller (19), with a precision of a few μrad and an adjustment time of approximately 0.3 ms (obtainable at Cambridge Technology,

Inc. MA, USA, model 6800).

The thus deflected bundle of beams traverses a Wollaston prism (12) which resolves the fluorescence light into two beams of orthogonal polarization (h: horizontally, and v: vertically polarized). The angle between the two rays is adjustable by the shape of the prism and here is chosen such that after having passed the imaging lens (13), the two polarization portions of the fluorescence light are imaged simultaneously but spatially separated on the pixel array, at the distance of half the x-width of the array. Such prisms are offered by a few companies, e.g. by Bernhard Halle Nachf. GmbH & Co., Berlin.

The combination of galvanometer mirror (11) and Wollaston prism (12) allows for an optimum utilization of the memory area of the pixel array, in synchronism with the frame transfer of the CCD camera for data shifting in the y-direction of the array. The utilization of this memory and data shift possibilities depends on the application thereof. Applications can be subdivided into three groups: either the focus plane is horizontally (in the x,y-plane) moved through the sample (SDT-x,y-scan), or vertically (3D-SDT), or both, horizontally (slowly) and vertically (rapidly and in cycles). These three modes of utilization of the same apparatus differ from each other only by their controller program, i.e. the control of the sample movement

(in x,y-direction or (and) in z-direction), in synchronism with the mode of illumination (wave lengths, light/dark intervals, etc.) and coordinated with suitable modes of data storage and shifting on the CCD pixel array (combined use of rotating mirror, Wollaston prism, frame transfer options and readout of the CCD camera).

Finally, the data collected with the CCD camera are processed, analyzed, and brought to be imaged by suitable means in an image generating unit (16).

The CCD camera which had been used for the estimates in Fig. 6 and Figs. 22, 23 plus description (see below) is obtainable at Princeton Instruments, USA, model MicroMAX-1300PB.

In the working mode for large-area screening (SDT-x,y-scan), the sample is continuously moved by precision motors (17) in x-direction (cf. Figs. 21 and 23).
Illumination is large-area (approximately 100 x 250μm) and continuous, so that each molecule will be illuminated for the same amount of time. At this mode, the rotating mirror remains in fixed angular position. The Wollaston prism may be used or pivoted out. The blind data on the CCD camera (cf. gray region in Fig. 23) are continuously shifted line by line in y-direction and read-out.

In the working mode for 3D imaging (real time 3D SDT) (cf. Figs. 14-18), individual fluorescence markers

on live cells can be imaged and their movement fol-
lowed.

In Fig. 20, the principle of 3D imaging is illus-
trated, whereas Fig. 21 with Fig. 22 show the signal-
logistic side of realization of the 3D imaging in terms
of apparatus, yet also merely for one example (two col-
ors with two polarizations each for each focus plane).
Other applications, e.g. excitation of three or even
more dyes, with or without polarization splitting, can
be realized by modifying the control program analogous
to the one described. Real time imaging of cells, ac-
cording to the criterion indicated below, for up to
four colors and registration of both polarization por-
tions are feasible without any problems.

As the CCD camera, in the present Example a Micro-
MAX-1300 PB with 1300 x 1340 pixels is used. The pixel
area is 20μm x 20μm. The frame shift is in y-direction
and requires 3 μs per line. Impacting photons are con-
verted into electrons by a conversion of 0.9 in the en-
tire optical range. For imaging times of seconds, the
camera is practically free from spontaneously forming
dark counts. Merely at a readout of the camera with 1
μs (20μs) per pixel, a weak background noise of 7 elec-
trons per pixel forms. As has been shown for the HASM
cell, approximately 150 electrons can be collected from
individual fluorophores in an illumination time of

merely 5 ms, which can be repeated a few times (3-20 times with the presently common dye molecules). When illuminating a 16 μm measurement area, the image size is approximately 80 x 80 pixels for a 100-objective. Fig. 22 shows the taking of 4 images for each z-position of the focus plane. The number in the single images gives the focus plane, rising from z=0 by 0.5 μm each.

The second index, r or g, stands for red or green fluorescence. On the first focus plane, at first red fluorescence is excited, which then is divided into both polarization portions (h and v), and in the left or right half, respectively, of the array, is collected in separate single images during the illumination time (cf. 1rh and 1rv). Subsequently, green fluorescence is excited and collected in the image areas 1gh and 1gv. To this end, the rotating mirror is rotated into the next position (approximately by 0.5 degrees) in the dark pauses of 0.5 ms each, so that both the "1gh" and the "1gv" image, shifted by 80 pixels towards x, fall onto the image region (gray), directly adjacent the two "red" images of level 1. In the following 0.5 ms dark pause, the focus level 2 is adjusted (by controlling the z-piezo on the objective), and the rotating mirror is turned into its new position, for the following imaging of 2rh and 2rv at first, and, after a mirror rotation, of 2gh and 2gv on this focus level 2. This is

continued until the image sequence (gray region) has been filled, in the present Example after the four imagings on level 4. In the subsequent pause of 0.5 ms, the 16 images are shifted by frame shift of the CCD camera in y-direction and 80 pixel lines, for which approximately 0.24 ms will be required, and the rotating mirror is moved back to its starting position. The image line marked in gray then is filled with further 16 images, from the focus planes 5 to 8, etc. When a predetermined z-value has been reached (exceeding the height of the imaged cell, approximately after 10µm, corresponding to 20 imaging planes), the 3D image is finished.

With an illumination time of 1.5 ms for each h and v image pair, each dye molecule, because of the overlap of the 1.6 µm deep focus plane, is effectively illuminated for approximately 5 ms, and its partial images in neighbouring planes contain approximately 150 electrons in the image pair.

The total image (20 planes) then merely requires 80 ms. During this time, membrane molecules (diffusion constants < 0.1 nm^2/s) or actively moved molecules (velocity < 2 µm/s) have not moved out of the optical resolution volume (ellipsoid with a length of half axis of 0.25 µm in x- and y-direction and 0.8 µm in z-direction). On the basis of this criterion, the method ac-

cording to the invention allows for producing 3D images of a complete cell in real time, and this with single molecule sensitivity, for molecular detection over the entire cell, a goal not reached before.

The read-out time of this 3D information requires approximately 1 second. Thus, every second a new 3D image can be captured, or at longer time intervals desired. In this manner, movements and processes of single molecules or associates of the entire molecule ensemble imaged can be observed (as long as the molecules are still fluorescent; under the illumination conditions indicated, at least 3 3D images can be made with the fluorophores presently common), yet also with a smaller and slower camera good results can be obtained for one dye and without polarization splitting.